

CODAN FRANCE SARL
49 RUE DE ROHRWILLER,
BP 40073
67242 BISCHWILLER CEDEX
REPUBLIC OF FRANCE

TEST CERTIFICATE

Product : 0.5 ML CODAN INS. U100 LUER. NEUTRAL GAS-PAPER
WITH FULL IN-LINE PRINT. INSERT IN BOX.

Quantity....: 19.200 **CODAN REF** 62.1912

LOT J85900-1 **Invoice No.....:** 447480

Steril.Date: 2024-02

Results

DS/EN ISO 11138

Biological Indicators for Ethylene Oxide Sterilization **Passed**

European Pharmacopoeia

2.6.14. Bacterial Endotoxins
(except oral tip one)

Passed

**The validation of the sterilization process
was performed according to:**

DS/EN ISO 11135-1

Sterilization of Health Care Products - Ethylene Oxide -
Requirements for Development, Validation and Routine Control
of a Sterilization Process for Medical Devices.

EN 556

Sterilization of Medical Devices -
Requirements for Medical Devices to be labelled "Sterile".

This certificate has been issued
electronically and is valid
without signature or stamp.

Date : Nov.18,24