



Certificate of Conformity

VWR Digital Timer, One Channel

VWR International certifies that the products comply in every respect with the relevant particular specifications and/or drawings and/or relevant standards and regulations in force, mentioned in this document.

VWR International certifie que les produits sont conformes à tous égards aux spécifications particulières appropriées et / ou dessins, normes et règlements en vigueur, mentionnées dans le présent document

VWR International bestätigt, dass das Produkte allen zutreffenden Spezifikations- und/oder Zeichnungs- und/oder Standardanforderungen sowie den im Dokument aufgeführten gültigen Vorschriften entspricht.

Catalog No.	Description	Colour	Accuracy	WxDxH (mm)	Pk
609-0206	Digital timer, single channel, 0.1 resolution	Violet/white	±0.01% (5 s)	81.3×71.1×30.5	1

Resolution : 0.1

Other Characteristics:

- Counts up and down plus displays elapsed time after countdown.
- Capacity of 99 minutes, 99 seconds.
- Audible and visual alarms.
- Last time set memory.
- Requires two AAA batteries.
- RoHS compliant.
- Individually tested and inspected using NIST in A2LA accredited laboratory (certification #2448.01), ensuring conformity to ISO 17025:2005.

This document has been produced electronically and is valid without a signature

Disclaimer: VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.

Date of issue: **October 2019**