

Certificate of Conformity

VWR Catalogue Number	See List
Description	VWR®, Cryo coder cap inserts
Country of Origin	Manufactured in India
Date of Issue (yyyy-mm-dd)	2024-02-05

VWR Catalogue Number	Colour
VWRI210-000033	Assorted
VWRI210-000034	Blue
VWRI210-000040	Green
VWRI210-000041	Grey
VWRI210-000036	Lilac
VWRI210-000035	Orange
VWRI210-000042	Pink
VWRI210-000037	Red
VWRI210-000038	Tan
VWRI210-000031	Violet
VWRI210-000032	White
VWRI210-000039	Yellow

Quality System Compliance

Products are manufactured under the **ISO 9001:2015 & ISO 13485:2016** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Test Name	Test Reference	Accuracy	Result
Autoclaving	QWI-QC-07	General Insp. Level-1, AQL -1.5	PASS
Dimensional	QWI-QC-23	General Insp. Level-1, AQL -1.5	Within tolerance

Product Specifications

Material	Polypropylene, 100% virgin, FDA 21 CFR, compliance, USP Class VI passed.
Sterilization	Not Applicable.
ATP Assay	Not Applicable.
DNase & RNase Free	The product is free of any detectable DNase/RNase contamination.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	The products are not made from natural rubber latex.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.
RoHS	Not Applicable.
REACH Statement	Not Applicable.
Storage Conditions	Store and transport at ambient temperature. Do not expose to extreme temperature and direct sunlight.
Shelf Life	Not Applicable.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoC using new layout	02-2024