HACH LANGE GmbH
Willstätterstr. 11
40549 Düsseldorf
Germany
Tel. + 49 (0) 211 5288-0
Fax. + 49 (0) 211 5288-210
info-de@hach.com
www.hach.com

DECLARATION OF CONFORMITY

I, the undersigned	:	**Bernd SEIDL** HACH LANGE GmbH Königsweg 10 D-14163 Berlin Germany
Position in the company	:	**Deputy Manager Calibration Lab.**
Declare that the products	:	**Code** 2283449.99 **Designation** pH 4.01 **Lot** 23038 **Expiration date** 08/31/2024

Comply with our specifications when sold:
Buffer solution : pH 4.01 ± 0.02 (25°C)

Our pH measurements are traceable to CRM (Certified Reference Material) from NIST (National Institute of Standards and Technology) and/or others national standards

Berlin, February 16, 2023

B. Seidl

Produced by:

HACH LANGE GmbH
Königsweg 10
14163 Berlin
Germany
www.hach.com