Kartell SpA DIVISIONE LABWARE	TECHNICAL DATA SHEET COD. 1052 Rev. 18/02/2022

TECHNICAL SPECIFICATIONS

Trade name	Sterilisation indicator tape
Use	Adhesive tape to indicate autoclaving sterilization To be used with individual units (i.e. packs, containers) to demonstrate that the unit has been exposed to the sterilization treatment and to distinguish between processed and unprocessed units. NOT INTENDED FOR USE AS VALIDATION SYSTEM OF STERILIZATION PROCESS.
Material (backing)	Impregnated semicreped paper with indicator strips
Type of adhesive	Natural rubber
Change that occurs after sterilization	Colour of the strips from white to dark brown
Minumum conditions for change	Steam 10 min 121°C or 2 min 134°C



DIMENSIONAL SPECIFICATIONS

Total thickness	mm. 0,15
Tensile strength	10,00%

GENERAL FEATURES

CE Mark	No
Conformity	ISO 11140/2015
Product classification	Indicators Class 1
Storage conditions	Temperature between 15 and 30°C - relative humidity between 30 and 70%
Storage time	36 months

PACKAGING

Minimum packaging	8 pcs single wrapped
Case	48 pcs single wrapped

Quality Assurance

This paper is an electronic print-out: signature is not requested