

CODAN FRANCE Sarl.
49 Rue de Rohrwiller
BP40073
67242 Bischwiller CEDEX
Französische Republik

*** TEST CERTIFICATE ***

Product: STERITEX® V-Airvent Air Venting Needle with check valve, sterile

Quantity	100	CODAN REF	64.6008
LOT	P84059-1	Invoice No	200914
Steril.Date	2019-11	Expiry Date	2024-10-28

Results

EN ISO 11135

Sterilization of health-care products

Passed

European Pharmacopoeia

The analysis of bacterial endotoxin was performed with the rFC test method to EP chapter 5.1.10

Passed

Date: 24 February 2021

This certificate has been produced electronically and is valid without signature.