

Certificate of Conformity for Riplate®, non-sterile

Article number: 43001-00621
Article name: RITTER RIPLATE 5 ML J.T.BACKER
Material: PP
Production mode: Injection molding
Lot no.: 995205/1
Production date: 07 2021
Expiry date: 07 2026

To assure that this product has passed all relevant quality inspections, the company Ritter is in accordance with and certified to the current valid ISO 9001, and has introduced and applied a quality system. Following these guidelines, the proof of fulfilment of the above-mentioned standard has been provided by the certification through TÜV Rheinland.

Ritter certifies that these plates are manufactured by state of the art processes, quality control procedures, raw materials, final inspections and meet all required specifications in accordance to test Riplate®. For special applications, appropriate methods and means, they must be verified and validated. An independent laboratory has tested the product to be free of human DNA, RNase, PCR inhibitors, ATP and with the following purity specifications:

Human DNA:

- Human DNA free => samples will show no evidence of human DNA with a detection level of < 0.5 pg/µl.

RNase:

- RNase free => samples will show no evidence of RNase with a detection level of < 1 x 10⁻⁹ KunitzU.

PCR inhibitors:

- PCR inhibitors free => samples will show no detection level.

ATP:

- ATP free => samples will show no evidence of ATP with a detection level of < 10⁻¹² mmol/µl.

Schwabmünchen, 2021-11-23

Stephan Schatz
COO