

Hiermit bestätigt die inpac Medizintechnik GmbH, dass nachfolgend gelistete Produkte / Kundenaufträge im Rahmen unten genannter inpac Bestellung sterilisiert wurden.

Herewith inpac Medizintechnik GmbH confirms the products / customer orders listed below were sterilized in the context of the inpac order mentioned below.

AUFTRAGSDATEN / ORDER DATA

Teilenummer / article No.	003037.000023
Teilebezeichnung / article	Verpackung diverse Tuben groß Fa. IKA-Werke
Kundenbestellnummer / PO No.	2500362
inpac Auftragsbestätigung / project No.	AB25/004059
Verpackte Stückzahl / pieces	710
Chargennummer / batch No.	2500546
Fertigungsauftragsnummer / production order No.	FL25/010001
Lieferscheinnummer / delivery receipt No.	LI26/000889
inpac Bestellnummer Sterilisation inpac order No. sterilization process	BD26/000033

BGS DATEN / BGS DATA

BGS Zertifikatsnummer / BGS certificate No.	1499925
BGS Auftragsnummer / BGS order No.	396685
BGS Lot Nummer / BGS batch No.	26010239/0
Zyklus / Cycle	GammaSteri BGS Bruchsal 6 Rund
Standort / Site	
Bestrahlungszertifikat / Sterilization certificate	i.O. / acceptable

Birkenfeld, 09/03/2026
Datum / Date

gez.: 1. Zitterbein
Name, Unterschrift / name, signature

Zertifikat



BGS John-Deere-Straße 1 + 3 76646 Bruchsal
Eurofins Inpac Medizintechnik GmbH
Gewerbestraße 34
75217 Birkenfeld

Your Delivery note No.: BD26/000033
Your Purchase No.: BD26/000033
Customer ID: 101443
Date: Mar 09, 2026

BGS Order No.: 396685

Certificate No.: 1499925
Page: 1 / 2

BGS-Lot-No.: 26010238/0

Irradiated quantity: 6 pallet(s), 64 carton(s)
Your article: Zyklus Medizinprodukte 7 Runden

BGS article number: 017223
Validation report: 017223-0-6-149923-E
Type of irradiation: gamma
Required minimum dose [kGy]: 25.00
Acceptable maximum dose [kGy]: 55.00
Irradiation date: Mar 08, 2026

Measured reference dose [kGy]: 40.82
Measured reference dose [kGy]: 41.46
Measured reference dose [kGy]: 40.02
Measured reference dose [kGy]: 41.62
Measured reference dose [kGy]: 40.98
Measured reference dose [kGy]: 40.02
Batch release: Follner Katja, Mar 09, 2026, 8:47 p.m.

BGS-Lot-No.: 26010239/0

Irradiated quantity: 1 pallet(s), 12 carton(s)
Your article: Zyklus Medizinprodukte 6 Runden

BGS article number: 040892
Validation report: 040892-0-5-169752-B
Type of irradiation: gamma
Required minimum dose [kGy]: 25.00
Acceptable maximum dose [kGy]: 50.00

