



Certificate of Compliance

Cole-Parmer Product Number: **95644-08**

Issued Date: **23 August 2022**

Lot#/Batch#: **29083293**

Your PO#: **N/A**

CP Order #: **N/A**

Description **SAMPLING SYSTEM 2L**

Compound: **N/A**

Specification: **N/A**

Cole-Parmer certifies that, according to the manufacturer, the product described meets the specifications set forth in this certificate, subject to the disclaimers set forth herein.

The C-Flex tubing used in this assembly meets the requirements of USP, Biological Test for Plastics, and Class VI.

This assembly has been gamma irradiated within the range of 25.0 kGy to 40.0 kGy.

This product is not approved or intended for, and should not be used for medical, clinical, surgical or other patient oriented applications.

It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Cole-Parmer certifies that, according to the most recent information provided by the manufacturer, the product described meets the specifications set forth in this certificate, subject to the disclaimers set forth herein.

This certification applies only to the condition of the above-described product in the unopened package. Cole-Parmer cannot and does not assume any responsibility for any failure of the product to meet this Certificate of Compliance that results from any handling or use subsequent to opening the package.

Cole-Parmer expressly warrants that the product listed above conforms only to the specifications herein, subject to the disclaimers listed above, if any. Except as otherwise set forth in writing by Cole-Parmer, there are no other warranties, express or implied, including without limitation warranties of marketability or fitness for a particular purpose, use or application, that apply to this certificate or the subject product. It is the responsibility of the user to determine the suitability of the subject product for any particular use or application.

Michael Whitaker

Regulatory Compliance Manager