# CERTIFICATE OF CONFORMANCE

| | |
|---|---|
| **Product Part Number:** | CSLB-12WHIR |
| **Lot Number:** | 026225-000 |
| **Irradiation Reference Number:** | 0005080-00 |
| **Lot Quantity:** | 4 CASE(S) |
| **Date of manufacture:** | 05/26/2021* |
| **Shelf Life:** | 24 MONTHS |
| **Description:** | WHITE MICROSEAL TAPE, IRR'D (72/CS) |
| **Size -** | |
| **Width:** | ½" |
| **Length:** | 18 YARDS |
| **Height:** | N/A |

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

06/10/2021

**Authorized Signature** **Date**
**Quality Control Department**

**CC:** **With Product**
**Customer File**
**C of C File**

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

© COPYRIGHT 2019
Form No. GEN-1034e
REV. No. 01

# STERIS: Gamma Certificate Of Processing

**Prepared for: MICRONOVA MFG INC (2727)**

**Gamma Process Run ID: 1113-3776A**

| Product Code | Lot Number | Quantity | UOM |
|---|---|---|---|
| CSLB-12WHIR | 026225-000 | 4 | Case |

**Processing Run Start Date: 08-Jun-2021 5:28 AM** **Approx. Downtime (hours): 2.48**

**Processing Run End Date: 08-Jun-2021 12:32 PM**

| | | | |
|---|---|---|---|
| **Minimum Specified Dose (kGy):** | **25.0** | **Minimum Delivered Dose (kGy):** | **28.5** |
| **Maximum Specified Dose (kGy):** | **40.0** | **Maximum Delivered Dose (kGy):** | **37.8** |

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

**Other Information**

0005080-00

## Signature Manifest

Reviewed and E-Signed By: **Joanna Jimenez (QS & RC Technician)**

Date/Time E-Signed: 10-Jun-2021 6:11 AM

Document Content Revision: 1

**Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.**

**Processing Location**

**1000 Sarah Place**
**Ontario CA 91761**
**United States**

Document ID: 80161

| Product Code | Lot Number | Quantitv | UOM |
|---|---|---|---|