


Test Report

– Eppendorf PCR clean –

EPENDORF TUBES® 5.0 mL



200 Tubes
PCR clean
centrifugation up to 25 000 x g

LOT K195438I

2026-03-28

Order no.: 0030 119.460
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030119460
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• Human DNA

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• PCR inhibition

Method: real-time PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 12.03.21

Testing of the above described lot showed conformity within the limits of detection.

16.03.21
Date of release

Epp 2103 8392
Registration number

J. V. P. Müller
Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1