

Certificate of Conformity

VWR Filter Tip, 10µl, Low Retention, Extra Long, Sterile

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 732-1614

Lot No: 007C4

Sterility assurance level: 10⁻⁶

Material: POLYPROPYLENE made of 100% virgin resin

Tested for heavy metals contamination using the U.S.P. (231) method and found to have results lower than 1ppm.

Properties: Free of detectable Nuclease contaminants as confirmed by gel assay

Free of Human DNA to a detection level of 1 pico-gram of DNA

It is tested and found to be free of Endotoxin to a detection level of 0.05 EU/mL

Certified ATP-free and Protease free

Certified free of Bisphenol A (BPA), phthalates and cytotoxic effects.

Bovine Spongiform Encephalopathy/ Transmissible Spongiform Encephalopathy (BSE/TSE) compliant.

Resins are USP Class-VI certified, RoHS, FDA regulation CFR 21, are free of Substances of Very High Concern (SVHC), REACH compliant, and is FDA approved for food contact.

The sterilization program is an ISO 11137 validated process, with bioburden studies and radiation validation audits performed quarterly.

This document has been produced electronically and is valid without a signature

Date of issue: **April 2021**