

Certificate of Conformity

VWR Transport Swab M40, Amies

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 710-0442

Lot No: 230783200

Expiry Date: 29th August 2024

Sterility: The applied dose range is validated to guarantee the Sterility Assurance Level SAL 10⁻⁶.
Sterilisation procedure validated in conformity with BS EN ISO 11137-1:2015 and BS EN ISO 11137-2:2015.

Goods manufactured according to: **ISO 13485:2016 system.**

Medium Visual Appearance and Mechanical Controls: **passed according to internal SOP.**

pH of the medium at 25 °C acceptability limits 7.2 ± 0.3: **passed**

Microscopic Bioburden Specifications, maximum 2 micro-organisms per 20 microscopic fields at 10X100 magnification: **passed.**

Cultural Response: swabs were inoculated with the tester organisms (see the table) and placed in the medium.
Recovery obtained after 24 hours at 20-25°C: **passed.**

Certificate of Conformity

TRANSYSTEM M40, AMIES AND STUART CLEAR GEL AND WITH CHARCOAL MEDIUM	TRANSYSTEM M40, AMIES AND STUART LIQUID MEDIUM
<i>Neisseria gonorrhoeae</i> ATCC 43069	<i>Neisseria gonorrhoeae</i> ATCC 43069
<i>Haemophilus influenzae</i> ATCC 10211	<i>Haemophilus influenzae</i> ATCC 10211
<i>Streptococcus Pyogenes</i> ATCC 19615	<i>Streptococcus pyogenes</i> ATCC 19615
<i>Streptococcus pneumoniae</i> ATCC 6305	<i>Streptococcus pneumoniae</i> ATCC 6305
<i>Bacteroides fragilis</i> ATCC 25285	/
<i>Propionibacterium acnes</i> ATCC 6919	/
<i>Peptostreptococcus anaerobius</i> ATCC 27337	/

This document has been produced electronically and is valid without a signature

Date of issue: **September 2023**