

CODAN FRANCE SARL
49 RUE DE ROHRWILLER,
BP 40073
67242 BISCHWILLER CEDEX
REPUBLIC OF FRANCE

TEST CERTIFICATE

Product : 3 ML CODAN 0.1 GRAD. LUER LOCK. NEUTRAL GAS-
PAPER WITH FULL IN-LINE PRINT. INSERT IN BOX.

Quantity...: 1.000 **CODAN REF:** 62.3760

LOT: H89289-1 **Invoice No.....:** 417242

Steril.Date: 2020-09

Results

DS/EN ISO 11138

Biological Indicators for Ethylene Oxide Sterilization **Passed**

European Pharmacopoeia

2.6.14. Bacterial Endotoxins
(except oral tip one)

Passed

The validation of the sterilization process
was performed according to:

DS/EN ISO 11135-1

Sterilization of Health Care Products - Ethylene Oxide -
Requirements for Development, Validation and Routine Control
of a Sterilization Process for Medical Devices.

EN 556

Sterilization of Medical Devices -
Requirements for Medical Devices to be labelled "Sterile".

Date : Apr.28,21

This certificate has been issued
electronically and is valid
without signature or stamp.