INSTALLATION GUIDE

RephiDuo U Pack 2（DI feed water）

RephiLe Catalogue Number:
RR400Q201
Corresponding Millipore Part Number:
QGARD00D2

Suitable Millipore Systems:
Milli-Q Academic, Biocel, Gradient, Synthesis, Element

This product is comparable to the corresponding Millipore cartridge. Please refer to Millipore's User Manual for a detailed instruction on installation.

A Brief description of the installation process is described below:

1. Put the system into " STANDBY" mode by pressing "OPERATE/STANBY" button for two seconds. **Do not turn off the electrical power.**
2. Open the point-of-use valve for a few seconds to depressurize the system.
3. Lift up the pack adapter cap.
   - Remove the metal locking clip from the metal guide pin.
   - Remove the old pack from the system.
4. Remove the two blue protective caps on the new cartridge. Wet the two O-rings on the cartridge with a few drops of pure water.
5. Push the new cartridge so that the pack adapter metal rod goes through the hole at the top of the cartridge.
   - Lift the cartridge and push the bottom of it into the small opening on the system.
   - Push the top of the cartridge towards the system until it is secure.
6. Lock the cartridge in place with the metal locking clip on the end of the metal guide pin.
7. Bring the adapter cap down to its lowest position so that it covers the top of the cartridge.

Then proceed to the installation of a new **RephiSolo U PAK.**



RephiSolo, RephiDuo and RephiQuatro are trademarks of RephiLe Bioscience, Ltd.
© 2010. RephiLe Bioscience, Ltd. All rights reserved.
All other trademarks are trademarks of their respective manufacturers.