



MEDICAL PRODUCTS - CLEVELAND FACILITY
911 North Davis Avenue, Cleveland, MS 38732, Phone: 662.843.9421, Fax: 662.846.5987

Certificate of Analysis

Product: 0.9% SODIUM CHLORIDE IRRIGATING SOLUTION

Lot #: G136937

Code: 2F7124

Expiry: NOV 2021

Manufactured Date: 11/09/2018

Sterilization Date: 11/09/2018

Mixing Specification: 03-15-19-156

Issue Date: 07/02/2002

TEST NUMBER	TEST NAME	LIMIT	RESULT
11-07-19-001	NaCl (g/L)	8.55 - 9.45 g/L	8.96 g/L
11-20-19-006	Sodium ID- USP	Positive	Positive
11-20-19-007	Sodium ID- Flame	Positive	Positive
11-21-16-003	pH at 25 deg. C	4.5 - 7.0	5.7
20-03-07-006	Particle Analysis - Large Volume (Instrumental)		
	Particles >= 10 micron	NMT 25 particles/mL >=10 microns	0 ct/mL
	Particles >= 25 micron	NMT 3 particles/mL >=25 microns	0 ct/mL
13-01-10-313	Bacterial Endotoxin	0.5 EU/mL	< 0.01 EU/mL
14-04-01-120	Sterility	Pass Parametric Release	Pass
PARAMETRIC RELEASE IN ACCORDANCE WITH U.S. FDA REQUIREMENTS			

THIS PRODUCT WAS MANUFACTURED ACCORDING TO CURRENT GMP,
FULLFILLS THE REQUIREMENTS OF THE MASTER PRODUCTION DOCUMENT AND TESTED TO USP COMPLIANCE

E. Soliz 11-13-18
Prepared By/Date

[Signature] 11-13-18
Quality Management Approval /Date

Ellen Soliz