

CERTIFICATE OF CONFORMANCE

Product Part Number: BL1-40TRIRBB
Lot Number: 023175-000
Irradiation Reference Number: 0004613-00
Lot Quantity: 20 CASE(S)
Date of manufacture: 01/28/2021*
Shelf Life: 36 MONTHS
Description: BINLINER, LDPE, 1 MIL, IRR'D (100/CS)
Size -
Width: 23"
Length: 17
Height: 48"

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

Sabito

02/08/2021

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared for: MICRONOVA MFG INC (2727)

Gamma Process Run ID: 1113-1825A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
BL1-40TRIRBB	023175-000	20	Case

Processing Run Start Date: 06-Feb-21 04:50 AM

Approx. Downtime (hours): 0.44

Processing Run End Date: 06-Feb-21 02:03 PM

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.2
Maximum Specified Dose (kGy):	40.0	Maximum Delivered Dose (kGy):	35.1

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0004613-00

Signature Manifest

Reviewed and E-Signed By: **Adrianna Garcia (QS & RC Technician I)**

Date/Time E-Signed: 2021-02-07 03:07 PM

Document Content Revision: 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place
Ontario CA 91761
United States