

Certificate of Conformity

VWR Catalogue Number	216-3103
Description	Tube Round Base 10mL PS Sterile
Country of Origin	Manufactured in Italy
Date of Issue (yyyy-mm-dd)	2024-06-19

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

The product is in conformance with the requirements of the following standards and regulations:

Material	Polystyrene
Sterilization	Products are sterilized using Beta rays and meets SAL 10 ⁻³ with a specified dosage range of 7.5 –60 kGy.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.

Latex Statement	The products are not made from natural rubber latex.
BPA Statement	Bisphenols are not used in the manufacture of the raw material and are not expected to be present.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	Not applicable.
Storage Conditions	Store away from sunlight, heat sources and humidity.
Shelf Life	60 months.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	06-2024