



Control No: 22QC-C-6588

Date: January 17, 2023

CERTIFICATE OF CONFORMITY

PRODUCT CODE	LOT NO.	STERILIZATION DATE	EXPIRY DATE
SR+OX2225C1	221215SC	2022-12	2027-11-30

I.	Visual Inspection	:	Conform
II.	Functional Inspection	:	Conform
III.	Sterilization	:	Conform
IV.	Bacterial Endotoxin Test	:	Conform

The visual and functional inspections are checked in accordance with Terumo specifications based on ISO standards.

Sterilization method : Electron Beam Sterilization

Sterilization dose is determined and audited periodically according to ISO11137.

Sterilization Assurance Level (SAL) is 10^{-6} .

The bacterial endotoxin limit for medical device is tested in accordance to USP standard.

TERUMO (PHILIPPINES) CORPORATION has established and applied a quality system for manufacturing of medical devices.

Report No.: 176125801-31

Registration No.: SX 2002969-1

Furnish requirement: EN ISO 13485:2016

Valid until: 2023-02-11

Maricris Arcilla
Quality Control Manager

TERUMO (PHILIPPINES) CORPORATION

124 East Main Ave. Laguna Technopark, Binan Laguna
Tel. No. +63 49 541 2111, Fax No. +63 49 541 2121