

Certificate of Conformity

VWR Latex Lab Glove, Powder-Free, Non-Sterile

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 112-5398

Lot No: 2021-07 L085121 2107

Size: S

Sampling plan: In accordance to ISO 2859-1: 1999

Parameter	Specification	Result	Sampling size
1.Dimension			
Palm width	85 ± 5	Min. 83 Max: 88	Pass
Length	Median : 240mm	Median : 249 mm	Pass
Cuff Thickness	Min. 0.09 mm	Min. 0.09 mm	Pass
Palm Thickness	Min. 0.12 mm	Min. 0.12 mm	Pass
Finger Thickness	Min. 0.13 mm	Min. 0.13 mm	Pass
2. Functionality and Appearance test			
Barrier	G1 AQL 1.5	Pass	
Visual	G1 AQL 2.5	Pass	

Certificate of Conformity

Parameter	Specification	Result	Sampling size
3. Lab testing			
Tensile Strength (Un-aged) (Mpa)	Min : 18 Mpa	19.0	Pass
Tensile Strength (Aged) (Mpa)	Min : 18 Mpa	18.8	Pass
Elongation at Break (Un-aged) (%)	Min : 650 %	800	Pass
Elongation at Break (Aged) (%)	Min : 500 %	720	Pass
Force at Break (After challenge test)(N)	Median : 6 N	6.8	Pass
Powder (mg)	≤ 2 mg	0.36	Pass

This document has been produced electronically and is valid without a signature

Date of issue: **December 2022**