

Certificate of Conformance and Irradiation

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code: BKTBAG2

Lot Number: 014700

Manufacturing Date: 12/10/2019

Expiration Date: 12/2021

Irradiation

Radiation Lot: 19121903

Irradiation Dose Specified: 25 kGy to 50 kGy

Irradiation Dose Delivered: 26.6 kGy to 36.8 kGy

Quality Department
Contec Cleanroom Technology (Suzhou) Co., Ltd.
2020-01-08



苏州中核华东辐照有限公司
Suzhou CNNC Huadong Radiation Co., Ltd.

文件编号 (Doc No.): CHF/QR-8.2.6-02

辐照证明书
CERTIFICATE OF IRRADIATION

执照登记号 (License No.): 913205097186218342 Serial No.: KT190080

客户名称 (Customer): 康泰无尘科技 (苏州) 有限公司

辐照日期 (Radiation Date): 2019/12/19 辐照加工批号 (Radiation Lot): 19121903

产品名称 Description	订单号 Order No.	产品货号 (代号) Stock No. (Code No.)	产品批号 Lot No.	箱数 Cases
无尘擦拭布	N/A	MEQT1002	014680	20
无尘擦拭布	N/A	LWCS0036	014682	19
无尘包装袋	N/A	BKTBAG2	014700	70
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

客户要求剂量 (Sterilization Dose Required by Customer):

最低剂量 (Minimum Dose): 25.0 kGy 最高剂量 (Maximum Dose): 50.0 kGy

苏州中核华东辐照有限公司特此证明上述产品已经钴-60射线辐照。

Suzhou CNNC Huadong Radiation Co., Ltd. hereby certifies that above products have been irradiated by cobalt-60 gamma ray.

剂量计名称 (Dosimeter Name): 重铬酸钾 (银) 剂量计批号 (Dosimeter Lot): 190202

剂量测量不确定度 (Dosimetry Uncertainty): 5%

产品实际最低吸收剂量 (Minimum absorbed dose): 26.6 kGy

产品实际最高吸收剂量 (Maximum absorbed dose): 36.8 kGy

签发 (Issued by): [Signature] 日期 (Date): 2019.12.20

复核 (Reviewed by): [Signature] 日期 (Date): 2019.12.20

本公司通过了ISO13485体系和ISO11137标准的认证。Our company is certified for conformance to ISO13485 and ISO11137 standards by TUV PS
地址: 江苏省吴江市松陵镇交通路4756号 Address: 4756# Jiaotong Road, Songling, Wujiang, Jiangsu, PRC