

CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

PRODUCT	LOT NUMBER	EXPIRY DATE	CERTIFICATION DATE
MW700-1	108129	31/12/2029	17/12/2024

Microloop™ Plastic Loop 10µl

PARAMETER	SPECIFICATION	RESULT
Sterilisation Method	Irradiation (Procedure conforms to ISO 11137)	PASSED
Visual inspection	Moulding must be completely smooth and without defect	PASSED

Product has been manufactured in accordance with ISO 13485:2016

Product has been manufactured in accordance with the requirements of the In Vitro Diagnostic Directive 98/79/EC 1998, Official Journal of European Communities, L331/1

For and on behalf of Medical Wire & Equipment



Quality Assurance