# Specification

## 1.07961.0100 Entellan™ new rapid mounting medium for microscopy

| | Specification | |
|---|---|---|
| Refractive index (n 20/D) | 1.490 - 1.500 | |
| Density (d 20 °C/ 4 °C) | 0.94 - 0.96 | |
| Fluorescence (as quinine at 365 nm) | ≤ 250 | ppb |
| Viscosity (20 °C) | 250 - 600 | mPa•s |
| Suitability for microscopy | passes test | |

Dr. Timm Schlegelmilch
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.