

☒ CERTIFICATE OF CONFORMITY

Part Number	Description	Quantity	Expiration Date	Lot #
200018SG	1ml TB Safety Syringe w/Fixed Needle 25G*5/8	500000 pcs	2026-12-08	01112051

No	Item	Criteria	Result		
			Pass	Fail	N/A
1	Biological Requirement	Per product, all parameters meet product specification and requirement	✓		
2	Chemical Requirement		✓		
3	Physical Requirement		✓		
4	Package Requirement		✓		
5	Sterility assurance: 10^{-6}		✓		

Sol-Millennium Medical, Inc. hereby certifies that the materials, processes and procedures used to manufacture the product indicated above are in compliance with the ISO13485-2016 and internal product specifications.

The product(s) indicated above were sterilized per ANSI/AAMI/ISO 11135:2014, Sterilization of health care products, Ethylene oxide: Requirements for development, validation and routine control of a sterilization process for medical devices and in compliance with FDA 21CFR 820.75 Process Validation. Each sterilization cycle is accompanied by strategically placed biological indicators that monitor the sterilization cycle and are tested in accordance with validation parameters established.

Name	Felix Zhang	Title	QARA Specialist
Signature	<i>Felix Zhang</i>	Date	2022-02-16