

Certificate Of Processing

Systagenix
Wound Management Ltd.
Gargrave, BD23 3RX, UK
Tel: +44 (0) 1756 747200

The products listed below have been irradiated at Gargrave Sterilisation Services, Systagenix, Gargrave, UK.

Customer: VWR International Ltd

Irradiation Batch Number/Schedule: 30,799

Irradiation Lot Start Date/Time: 08-Dec-21 7:00:09AM

Irradiation Lot End Date/Time: 08-Dec-21 1:05:41PM

Number Of Interruptions: 1

Comment: Product Description: Ergo 500ml Sq

PO: 4514044908

Expiry: 28/10/2023

Inventory Detail

Finished Goods Code	Lot Number	Released Quantity	Delivered Dose (kGy)		Specified Dose Range (kGy)
			Minimum	Maximum	
VWRI331-0280	500BG/74	40	20.0	29.2	15.0 - 40.0

Gargrave Sterilisation Services/Systagenix certifies that the materials listed above received the absorbed doses within the precision and accuracy of the dosimetry system employed.

Signature Manifest

Signed By Supervisor **Andrew Stanners**

Signed On: 09-Dec-21 At: 9:20 AM GMT:0:00

Signed By QA **Sharon Allport**

Signed On: 09-Dec-21 At: 11:02 AM GMT:0:00

Document Content Revision: 1