



## CERTIFICATE OF CONFORMANCE

**Product Part Number:** T6-14IR  
**Lot Number:** 049988-000  
**Irradiation Reference Number:** 0010701-00  
**Lot Quantity:** 48 CASE(S)  
**Date of manufacture:** 06/10/2025\*  
**Shelf Life:** 60 MONTHS  
**Description:** TMOP COVER, NOVAPOLY, IRR'D (12/CS)  
**Size -**  
**Width:** 2.5"  
**Length:** 14"  
**Height:** N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

6/19/2025

**Authorized Signature**  
**Quality Control Department**

**Date**

**CC:** With Product  
Customer File  
C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)  
Gamma Process Run ID 1113-26324A

| <u>Product Code</u> | <u>Lot Number</u> | <u>Quantity</u> | <u>UOM</u> |
|---------------------|-------------------|-----------------|------------|
| T6-14IR             | 049988-000        | 48              | Case       |

Processing Run Start Date 17-Jun-2025 11:51 PM  
Processing Run End Date 18-Jun-2025 6:49 AM

|                              |      |                              |      |
|------------------------------|------|------------------------------|------|
| Minimum Specified Dose (kGy) | 25.0 | Minimum Delivered Dose (kGy) | 26.2 |
| Maximum Specified Dose (kGy) | 40.0 | Maximum Delivered Dose (kGy) | 36.5 |

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information  
0010701-00

Signature Manifest

Reviewed and E-Signed By: **Tricia Nguyen (Quality Analyst)**  
DateTime Esigned 19-Jun-2025 10:17 AM  
  
Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location  
1000 Sarah Place  
Ontario CA 91761  
United States

**APPROVED**  
*By Jenicah Baquir at 10:21 am, Jun 19, 2025*