

Certificate of Conformity

VWR Forcep

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 232-0192

Lot No: 20210701

Date of Manufacture: 01st Jul 2021

Material: 100% High Clarity Polystyrene

Length: 12.5 cm

Quality System Compliance:

Products are manufactured under the EN ISO 13485:2016 standard. Products are inspected and controlled through the whole production process in accordance with current applicable product specifications and QC SOP. Inspection is performed according to AQL level S-2 1.5 sampling plan per ISO 2859-1:1999. Inspection records are reviewed and signed off by qualified personnel for product release.

Product Specifications:

This product is a Class I category device as defined by the FDA in 21CFR Parts 862-892.

BSE/TSE Statement: The above product contains no, or has not been in contact with, any ingredients of animal origin and/or material derived from or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/Bovine Spongiform Encephalopathy (BSE).

Certificate of Conformity

Non-Pyrogenic: N/A

DNase & RNase Free: N/A

Cytotoxicity: N/A

Latex Free Statement: N/A

This document has been produced electronically and is valid without a signature

Date of Issue: **March 2022**