




CERTIFICATE OF CONFORMANCE

Product Part Number: PEN-01IR
Lot Number: 014642-000
Irradiation Reference Number: 0002668-00
Lot Quantity: 40 CASE(S)
Date of manufacture: 07/28/2019
Shelf Life: N/A
Description: CLEANROOM PEN, BLACK, IRR'D (100/CS)
Size -
Width: N/A
Length: N/A
Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.



Authorized Signature
Quality Control Department

07/31/2019

Date

CC: With Product
Customer File
C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel (310) 784-6990 • FAX (310) 784-6980

Certificate Of Processing

Prepared for MICRONOVA MFG INC



Gamma Process Run ID 116734A

Product Code

Product Lot Number

Quantity

UOM

PEN-01HR

014642-000

40

CS

Other Information:

P.O. #0002668-00

Processing Run Start Date/Time:	28-Jul-2019 10:01:17 am	Approx. Downtime (hours):	0.37
Processing Run End Date/Time:	28-Jul-2019 03:19:06 pm		

Minimum Specified Dose (kGy):	40.0	Minimum Delivered Dose (kGy):	46.1
Maximum Specified Dose (kGy):	60.0	Maximum Delivered Dose (kGy):	58.9

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By

 **Camargo, Denise (QSRC Technician I)**

Document Content Revision: 1

Signed On 7/30/2019 at 10:10 AM

UTC / GMT Offset (hh:mm): -7:00

Processing Location:

STERIS
1000 South Sarah Place
Ontario, CA 91761
Phone: 909-390-9942
Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.