Certificate of Sterility

Article Number: 300-0232 **VWR Sterile transport bag and surface sampling sponge with handle pre-moistened with Neutralising Buffer**

Lot Number: J098303

Expiry Date: 08/2024

VWR International hereby certifies that the specified meet all VWR specifications and quality criteria for release.

- All Surface samplers, sponge manufactured in accordance with HACCP, USDA, ACIA and CFIA requirements. Specifications: biocide-free cellulose sponge maintains viability of organisms.
- Sterilized by gamma irradiation and packed in sterile sampling bag with leakproof closure system accompanied by a foil sealed, pre-moistened cellulose sponge, holding 10 ml of the buffer.
- Random tests carried out for Total Plate Count for Aerobic and Anaerobic Bacteria, Yeast and Molds, based on USP 40 section 71 modified testing methods which indicated no bacteria or molds and therefore VWR confirms the product to be sterile.

This document has been produced electronically and is valid without a signature

Date of issue: **March 2024**