

Certificate of Compliance

This is to certify:

Part #: V1545C-FM

Description: 15X45 Clear Vial with 13-425 Foam Lined Cap

Lot #: 3-297-001

Clear glass vial made of Type 1 Borosilicate glass conforms to the following specifications:

Specific Vial Size: 15mm x 45mm

Tubing OD: 15.50mm (+/- .25mm) Clear Type 1 Borosilicate Tubing

- The glass meets the requirements for Type I glass as set forth in the current edition of the U.S. Pharmacopoeia (USP) section <660>, Glass Grains Test.
- The glass has been tested and is in compliance with specifications set forth in the current revision of ASTM 438-92 Specification for Glasses in Laboratory Apparatus.
- The glass has been tested and is in compliance with hydrolytic resistance of powder glass specifications as set forth in the current edition of the European Pharmacopoeia (EP) under section 3.2.1 Test B.
- This glass will meet the Soluble Alkali Test requirements for (containers to be fused) and for (containers not to be fused), as set forth in the current edition of the Japanese Pharmacopoeia Chapter 7, Tests for Glass Containers for Injection 7.01.
- This glass will meet the arsenic extraction requirements as specified in the current revisions of USP and EP headings referenced above.
- This glass meets the heavy metal requirements set forth in EU Packaging and Packaging Waste Directive 94/62/EC and in the CONEG Toxics in Packaging regulation.