

**Certificate of Analysis**

**Product: 0.9% SODIUM CHLORIDE IRRIGATING SOLUTION**

**Lot #: G105148**  
**Code: 2F7124**  
**Expiry: NOV/2015**

**Manufactured Date: 11/26/2012**  
**Sterilization Date: 11/26/2012**

| TEST                | LIMIT                                 | RESULT   |
|---------------------|---------------------------------------|----------|
| NaCl (g/L)          | 8.55 - 9.45 g/L                       | 8.91 g/L |
| Sodium ID- Flame    | Positive                              | Positive |
| Sodium ID- USP      | Positive                              | Positive |
| pH at 25 deg. C     | 4.5 - 7.0                             | 5.6      |
| Particle Analysis   | NMT 25 particles/mL $\geq$ 10 microns | 0        |
| Particle Analysis   | NMT 3 particles/mL $\geq$ 25 microns  | 0        |
| Bacterial Endotoxin | < 0.25 Eu/mL = Pass                   | Pass     |
| Sterility           | Pass Parametric Release               | Pass     |

PARAMETRIC RELEASE IN ACCORDANCE WITH US. FDA REQUIREMENTS

THIS PRODUCT WAS MANUFACTURED ACCORDING TO CURRENT GMP, FULLFILLS THE REQUIREMENTS OF THE MASTER PRODUCTION DOCUMENT AND TESTED TO USP COMPLIANCE.

E. Soliz 11-28-12  
Prepared By/Date

ELLEN SOLIZ

Twila Proctor 11-30-12  
Quality Management Approval /Date

Twila Proctor