



Certificate of Conformance, Analysis, and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PSQS0001
Lot Number:	731950-0-0-1
Manufacturing Date:	10/08/2019
Isopropyl Alcohol:	70 ± 3% by volume with the remainder DI H2O
Expiration Date:	10/2021

Analysis

Quantity of Particles Po:	3.29 millions/m ²
Quantity of Fibers Po:	0.034 thousands/m ²
Sorbency Rate:	1.0 seconds
Nonvolatile residue in DI:	0.00 g/m ²
Nonvolatile residue in IPA:	0.00 g/m ²

Sterility

Gamma Processing Run ID:	2532975
Irradiation Dose Specified:	25 kGy to 50 kGy
Irradiation Dose Delivered:	28.4 kGy to 41.2 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

QC Supervisor
coc@contecinc.com

Wednesday, October 16, 2019

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec, Inc.
ZI du Prat - R.P. 3707
F-56037 VANNES Cedex France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 2201B

Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
 TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

10/14/2019 14:58:48 GMT
 Page 1 of 1

Customer Name:	Contec- PO Box 530	Processing Facility:	Charlotte Gamma	Work Order #	2532975
P.O.#	103978			Sales Order #	2343142
25-50 kGy		CNT1, Gamma Treatment	Irradiation Date/Time:	10/11/2019 21:51:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	108	CA	89499-420	STERILE 70% IPA/ 30% DI 32 OZ	727148-0-0-1	103978
102.000	54	CA	PSPS0076	Clean Room Product	731927-0-0-1	103978
103.000	36	CA	PSQS0001	Gamma Wipes	731950-0-0-1	103978
104.000	24	CA	PSWE0001	Clean Room Products	723694-0-0-1	103978
105.000	108	CA	SB327030IR	32 oz IPA Spray Bottle 70/30	726354-0-0-1	103978
106.000	108	CA	SB327030IR	32 oz IPA Spray Bottle 70/30	726355-0-0-1	103978
	438	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	28.4 KGY	Pass	LSNIDER	10/12/2019 09:00:02 GMT
	Reason Code Test					Lonnie Snider	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	41.2 KGY	Pass	LSNIDER	10/12/2019 09:00:19 GMT
	Reason Code Test					Lonnie Snider	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
 Reason: Work Order Completions

Date: 10/14/2019 14:56:25 GMT