

Certificate of Conformity

Product Code: LW4010

Item Description: 1ml Graduated Pastette Sterile 20s

Lot Number: 25490587

Material: Low Density Polyethylene

Sterility: Sterile

Sterilisation Method: E-beam Irradiation

Sterility Details: Sterility Assurance Level: 10^{-6}

Where a Sterility Assurance Level is stated, it is ensured by irradiating the product according to ANSI/AAMI/ISO 11137-2015 guidelines.

Date of Manufacture: 13-DEC-2025

Date of Expiry: 01-DEC-2029

Manufacturing Location: Mexico

Other information relating to this product:

- Length: 155mm
- Bulb Draw: 3.4ml
- Drops per ml: 25
- Drop Volume: 40µl
- Graduations at 0.25ml

Alpha Laboratories Ltd hereby certifies that these products have been manufactured in a controlled environment in accordance with established manufacturing guidelines. Product inspection is achieved using appropriate AQL Levels per product approved requirements. All items have been inspected and meet packaging requisites.

Certified by:



Michelle Holland, Managing Director

Date: 24 July 2026



Alpha Laboratories Ltd
40 Parham Drive
Hampshire SO50 4NU

Tel: +44 (0)23 8048 3000
Email: alpha@alphalabs.co.uk
Web: www.alphalabs.co.uk