



Certificate of Conformity

VWR Sampling Scoop, 150 ml, White

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.: 231-0491

Lot.No: 200603BUR53789902

Material: Polystyrene

Length: 216 mm

Type of Irradiation: Gamma

Date of Irradiation: 03rd July 2020

Required Minimum Dose: 17.5 kGy

Determined Minimum Dose: 22.0 kGy

Determined Maximum Dose: 29.5 kGy

This meets the requirements of FDA 21 CFR part 11.

Expiry Date: June 2022

This document has been produced electronically and is valid without a signature

Date of issue: **November 2020**