



Technical Datasheet

VWR VINYL GLOVES DISPOSABLE, NON STERILE, POWDERED

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, dass die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Intended Use:

General uses
Food processing

Specifications:

Material	Polyvinylchlorid (PVC)	
Shape	Ambidextrous	
Color	Clear	
Inner	Pre-powdered	
Outer	No treatment	
Thickness (double measured, minimum values)	Finger	min. 0.10 mm
	Palm	min. 0.14 mm
	Cuff	min. 0.08 mm
AQL (acc EN 420)	Level 2 – AQL 1.5	
Residual powder (according to EN 455-3)	<15mg/dm ²	
Cuff/ Surface	Rolled cuff/Smooth	
Length (according to EN 455-2)	Median 240 mm	
Force at break (according to EN 455-2)	Median 3.6 N (during shelf life)	



Technical Datasheet

VWR VINYL GLOVES DISPOSABLE, NON STERILE, POWDERED

Packaging:

Size	Catalogue Number	Gloves/Dispenser box	Dispenser box / case
Small (6-7)	112-2758	100	10
Medium (7-8)	112-2759	100	10
Large (8-9)	112-2760	100	10
X-Large (9-10)	112-2771	90	10
Best storage conditions	10° - 30°		
	60% - 80% humidity		
	Dark & dry		
Shelf Life	5 years		
Country of Origin	China		

Markings:



Compliance to European Directives and Regulations:

PPE Regulation (EU) 2016/425, Personal protective equipment, CAT I
Council Directive 1993/42/EC amended by 2007/47/EC, relating to Medical Devices, Class 1
Regulation EC 1935/2004 - Regulation on materials and articles intended to come into contact with food

Technical Datasheet

VWR VINYL GLOVES DISPOSABLE, NON STERILE, POWDERED

European Standards:

EN 420:2003+A1:2009 - Protective gloves – General Requirements and test methods
EN 455-1:2000 Medical gloves for single use Part 1: Requirements and testing for freedom from holes
EN 455-2:2015 Medical gloves for single use Part 2: Requirements and testing for physical properties
EN 455-3:2015 Medical gloves for single use Part 3: Requirements and testing for biological evaluation
EN455-4: 2009 – Medical Gloves for single use Part 4: Requirements and testing for shelf life determination

Caution:

Vinyl gloves are not suitable for fatty food!
--

Disclaimer: VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.