

Certificate of Conformity

VWR Catalogue Number	216-1161
Description	Vacuum Urine Containers, 120 mL Capacity, Sterile
Lot Number	23C0685
Expiry Date (yyyy-mm)	2028-02
Country of Origin	Manufactured in Italy
Date of Issue (yyyy-mm-dd)	2026-03-13

Quality System Compliance

This certificate confirms that the product has been sourced from a supplier or manufacturer who has declared their Quality Management System (QMS) complies with **ISO 13485:2016**.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

These products are in compliance with CE IVD Directive 98/79/EC.

Material	Cup – Polypropylene. Cap – Polyethylene.
Sterilization	Products are sterilized using Beta rays and meets SAL 10 ⁻⁶ with a specified dose range of 17.8-20.4 kGy.
ATP Assay	Not tested.
DNase & RNase Free	Not tested.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not tested.
Non-Pyrogenic Statement	Not tested.
Latex Statement	This product is not made from natural rubber latex.
BPA Statement	Bisphenols is not used in the manufacture of the raw material and are not expected to be present
DEHP Statement	DEHP has not been intentionally incorporated in the manufacturing of the product.
RoHS	Not Applicable.
REACH Statement	Not Applicable.
Storage Conditions	Store in a dry place, away from sunlight.
Shelf Life	60 months.

Disclaimer: VWR's Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability, and distribution processes. VWR is not a manufacture and does not engage in manufacturing, co-manufacturing or processing activities such as design, packaging, labeling, testing or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use.

This document has been produced electronically and is valid without a signature.