

Certificate of Quality



Part Number	MFLX19406-16
Lot Number	0223755578
Part Description	SENSOR TEMPERATURE PS 3/8IN HB
Expiration Date	01JUN2024
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: Polysulfone
- Product is manufactured in a Class 5 clean environment.
- All plastic materials in contact with product fluid path meet USP Class VI requirements.
- Product is Animal Derived Component Free.
- Each thermistor is tested at 10°C to confirm 4482# $\pm$ 0.1°C.
- Each thermistor is tested at 25°C to confirm 2252# $\pm$ 0.1°C.
- Each thermistor is tested at 40°C to confirm 1200# $\pm$ 0.1°C.
- Each product is tested to confirm 500VDC isolation minimum between thermistor assembly and stainless steel tube.
- Each product is tested at room temperature to ensure proper assembly and electrical continuity.
- This product is not approved or intended for, and should not be used for medical, clinical, surgical or other patient oriented applications.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.