



CERTIFICATE OF MANUFACTURE/ANALYSIS

Product Size:	<u>200 Gallons</u>	Product Name:	<u>ProKlenz Booster</u>
Manufacture Date:	<u>April 4, 2019</u>	Product ID:	<u>1609-D9</u>
Exp. Date:	<u>April 1, 2022</u>	Lot No.:	<u>299905</u>

I. The product listed above is made per the current issue of the following documentation:

Raw Material Quality Control Specifications
Finished Product Master Formula
Manufacturing Instructions
Packaging Specifications
Finished Product Quality Control Specifications

Copies of this documentation are on file at STERIS Corporation, St. Louis Operations and will be provided as required to the applicable regulatory agencies.

II. The product listed above has been analyzed. The release specifications and results are shown below.

<u>Test Name</u>	<u>Method No.</u>	<u>Results</u>	<u>Specification</u>
Color	110	Straw	Colorless to Straw
Clarity	107	Clear	Clear
pH	106	5.6	4.5 – 6.0
Hydrogen Peroxide Assay	252	5.1%	4.5 – 5.5%
Specific Gravity (@ 25°C)	114	1.039	1.035 – 1.045
Analyzable Surfactant Assay	LC23	5.1%	4.5 – 5.5%

Based on these results the product is released for shipment to all STERIS Corporation distribution points.

Signed: *E. Eason*

Date: April 9, 2019

Quality Representative