



LP ITALIANA SPA

Via C. Reale 15/4, 20157 Milano, ITALY
T +39 02 3933 06.1 - F +39 02 3931 3484
www.lpitaliana.com - info@lpitaliana.com

Capitale Sociale € 309.600,00
R.E.A. MI 882798
Reg. Imp. C.F. e P.I. MI 01794050151

DECLARATION OF CONFORMITY Sterile Test Tubes

References: Invoice FV-21-00573 of FEB. 26, 2021

Product Code	Description	Lot n.
111018	PS STERILE CYLIND. 10 ML TUBE, 16X100 MM	A0249AAA
		Expiry Date
		2026-02

LP ITALIANA declares that all quality qualifications of the products have been respected and that all information on these documents are correct.

- Products are manufactured in accordance with quality system ISO 13485 and ISO 9001 ☒
- Products passed visual and functional controls and are in accordance with our internal procedures. ☒
- Irradiating process is validated and in accordance with ISO 11137 ☒
- Products are irradiated by ionizing radiations at nominal dose of 18 kGy. Batch number 4334 ☒

LP ITALIANA SPA

Massimiliano Capitanio

Quality Assurance Manager