

CODAN Medical ApS · Færgevej 4 · DK-4970 Rødby

CODAN Medical ApS

Telefon +45 54 65 74 11
www.codancompanies.com
CVR-nr.: 16 27 20 78

CODAN FRANCE SARL
49 RUE DE ROHRWILLER,
BP 40073
67242 BISCHWILLER CEDEX
REPUBLIC OF FRANCE

TEST CERTIFICATE

Product : 1 ML CODAN U100 LUER. NEUTRAL GAS-PAPER WITH
FULL IN-LINE PRINT. INSERT IN BOX.

Quantity....: 3.100 **CODAN REF** 62.1640

LOT J81828-1 **Invoice No.....:** 427174

Steril.Date: 2021-11

Results

DS/EN ISO 11138

Biological Indicators for Ethylene Oxide Sterilization **Passed**

European Pharmacopoeia

2.6.14. Bacterial Endotoxins
(except oral tip one)

Passed

**The validation of the sterilization process
was performed according to:**

DS/EN ISO 11135-1

Sterilization of Health Care Products - Ethylene Oxide -
Requirements for Development, Validation and Routine Control
of a Sterilization Process for Medical Devices.

EN 556

Sterilization of Medical Devices -
Requirements for Medical Devices to be labelled "Sterile".

Date : Mar.16,22

This certificate has been issued
electronically and is valid
without signature or stamp.