



Certificate of Conformance and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PSC20001
Lot Number:	697497-0-0-1
Manufacturing Date:	04/17/2019
Expiration Date:	04/2021

Sterility

Gamma Processing Run ID:	2423270
Irradiation Dose Specified:	25 kGy to 50 kGy
Irradiation Dose Delivered:	29.6 kGy to 39.5 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

A handwritten signature in black ink, appearing to read "C. Duffee".

QC Supervisor
coc@contecinc.com

Tuesday, April 23, 2019

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec, Inc.
ZI du Prat - R.P. 3707
F-56037 VANNES Cedex France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

Page 1 of 1
web: www.contecinc.com
e-mail: info@contecinc.com

Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

04/22/2019 12:44:17 GMT
Page 1 of 1

Customer Name:	Contec- PO Box 530	Processing Facility:	Charlotte Gamma	Work Order #	2423270
P.O.#	99772			Sales Order #	2216652
	25-50 kGy	CNT1, Gamma Treatment	Irradiation Date/Time:	04/20/2019 13:12:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	142	CA	PSC20001	Clean Room Product	697497-0-0-1	99772
	142	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	29.6 KGY	Pass	MFREIBERG	04/21/2019 13:23:07 GMT
		Reason Code Test				MICHAEL FREIBERG	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	39.5 KGY	Pass	MFREIBERG	04/21/2019 13:23:18 GMT
		Reason Code Test				MICHAEL FREIBERG	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
Reason: Work Order Completions

Date: 04/22/2019 12:42:34 GMT