

# Certificate of Conformity

|                                   |                                 |
|-----------------------------------|---------------------------------|
| <b>VWR Catalogue Number</b>       | See List                        |
| <b>Description</b>                | Single-Well Cell Culture Plates |
| <b>Country of Origin</b>          | Manufactured in China           |
| <b>Date of Issue (yyyy-mm-dd)</b> | 2022-11-18                      |

| Catalogue Number | Description                    | No. Of Wells | Culture area [cm <sup>2</sup> ] | Recommended working volume [mL] | Well volume [mL] |
|------------------|--------------------------------|--------------|---------------------------------|---------------------------------|------------------|
| 734-2977         | Single-well plate, non treated | 1            | 97                              | 10 - 40                         | 73               |
| 734-2966         | Single-well plate, TC treated  | 1            | 97                              | 10 - 40                         | 73               |

## Quality System Compliance

Products are manufactured under the **ISO 9001:2015 & ISO 13485:2016** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

## QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

## Product Specifications

The product is Class I category device as defined by FDA in 21CFR Parts 862-892.

|                          |                                                                                                                                                                                                                                                                               |
|--------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Material</b>          | Polystyrene (PS), Compliance with USP Class VI.                                                                                                                                                                                                                               |
| <b>Temperature range</b> | -20 to +50 °C.                                                                                                                                                                                                                                                                |
| <b>Sterilization</b>     | Product labeled as sterile is EB (Electron beam) irradiated and dose released upon ISO11137 recommended practices in effect at the time of validation. Products labeled sterile meets a minimum requirement of 10 <sup>-6</sup> SAL with a specified dose range of 15-30 kGy. |
| <b>ATP Assay</b>         | Not Applicable.                                                                                                                                                                                                                                                               |

|                                |                                                                                                                                                                                                                                                                                                                                                                                                 |
|--------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>DNase &amp; RNase Free</b>  | This product is free of any detectable DNase/RNase contamination.                                                                                                                                                                                                                                                                                                                               |
| <b>BSE/TSE</b>                 | No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE) according to EMEA 410/01 revision 3. |
| <b>Cytotoxicity</b>            | Testing is conducted to quality all material resins using ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.                                                                                                                                                                                                                                                             |
| <b>Non-Pyrogenic Statement</b> | The acceptance level for product is 0.5 EU/ml or less than 20 EU/device. (TAL Gel Clot Method). Test passed.                                                                                                                                                                                                                                                                                    |
| <b>Latex Statement</b>         | This product is not made from natural rubber latex.                                                                                                                                                                                                                                                                                                                                             |
| <b>BPA Statement</b>           | Bisphenol are not used in the manufacture of the raw material and are not expected to be present.                                                                                                                                                                                                                                                                                               |
| <b>DEHP Statement</b>          | Not Applicable.                                                                                                                                                                                                                                                                                                                                                                                 |
| <b>RoHS</b>                    | No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed).                           |
| <b>REACH Statement</b>         | Not Applicable.                                                                                                                                                                                                                                                                                                                                                                                 |
| <b>Storage Conditions</b>      | Room Temperature.                                                                                                                                                                                                                                                                                                                                                                               |
| <b>Shelf Life</b>              | 3 years.                                                                                                                                                                                                                                                                                                                                                                                        |

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.  
This document has been produced electronically and is valid without a signature.