

Certificate of Conformity

VWR Catalogue Number	612-3823
Description	Volumetric flasks, Class A, amber glass, with Polyethylene Stopper, 20 mL Capacity, 10/19 NS Socket Size, 0.04 ± ml Tolerance
Lot Number	GL0320
Date of Manufacture (yyyy-mm-dd)	2020-01-13
Country of Origin	Manufactured in India
Date of Issue (yyyy-mm-dd)	2022-07-21

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard and calibrated in **ISO 17025:2017** certified Laboratory. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

The product is in conformance with the requirements of the following standards and regulations:

Technical Standard for Products: EN ISO 1042.
Calibration Procedure: ISO 4787.

Material	Borosilicate 3.3 Glass DIN ISO 3585:1998. TYPE-1. CLASS-A. and meets ASTM E-438. TYPE-1. CLASS-A Standard.
Sterilization	Hot air sterilization in the oven up to 140 °C Steam sterilization in an autoclave 121 °C/ 20 min 134 °C/ 10 min.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	Not applicable.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2022-01-17 are used in manufacturing of the raw materials used to produce this product. No routinely analysis is performed.
Storage Conditions	Stored at room temperature in normal warehouse condition.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature