

Certificate of Conformity

VWR Catalogue Number	620-0897
Description	Precision thermometers, -1...+101 °C range, ±1.0 °C accuracy, 0.1°C division, 610 mm length, 76 mm immersion depth
Lot Number	5867
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2023-12-20

Quality System Compliance

Products are manufactured under the **ISO 9001:2015 and ISO/IEC 17025:2017** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Lead-free Glass, thermometer with Blue Organic Liquid Fill.
Sterilization	Non-Sterile.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.
BSE/TSE	Not applicable.
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.

RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2023-06-14 are used in manufacturing of the raw materials used to produce this product. No routinely analyse is performed.
Storage Conditions	Store at Ambient warehouse temperatures, low humidity, avoid exposure to direct sunlight.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.