



CERTIFICATE OF CONFORMANCE

Product Part Number: BL1-40TRIRBB
Lot Number: 039632-000
Irradiation Reference Number: 0007802-00
Lot Quantity: 20 CASE(S)
Date of manufacture: 08/07/2023*
Shelf Life: 36 MONTHS
Description: BINLINER, LDPE, 1 MIL, IRR'D (100/CS)
Size -
Width: 23"
Length: 17
Height: 48"

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

8/14/2023

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared for: MICRONOVA MFG INC (2727)
Gamma Process Run ID 1113-16294A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
BL1-40TRIRBB	039632-000	20	Case

Processing Run Start Date: 12-Aug-2023 9:42 PM Approx. Downtime (hours): 0.53
Processing Run End Date: 13-Aug-2023 3:47 AM

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.6
Maximum Specified Dose (kGy):	40.0	Maximum Delivered Dose (kGy):	36.6

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information
0007802-00

Gamma Process Run Approval authorized by STERIS
Date/Time: 13-Aug-2023 5:06 PM

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location
1000 Sarah Place
Ontario CA 91761
United States

APPROVED
By Noel Chaidez at 10:10 am, Aug 14, 2023