



L30357010

OKC Allergy Supplies, Inc.

a division of

ALK-Abello, Inc.

Certificate of Analysis

Product: SEV - Sterile Empty Vial
Lot number: SEV2031220
Vial size: 20mL-20mm
Manufacture date: 3/12/2020
Part Number: SEV20

Quality Release Testing		
Assay	Specifications	Results
Appearance - Product	Essentially free of visible particulate	Pass
Appearance - Packaging	No apparent leakage or physical alteration	Pass
Sterility, USP <71>	Sterile	Conforms
Bacterial Endotoxins, USP <85>	≤ 0.25 EU/mL	≤ 0.0050 EU/mL

Release Date	Expiration Date
4/22/2020	03 - 2024

Quality Assurance Approval	
By / Date:	<i>Cheryl E. Eschler</i> 04-27-2020

Original Certificate