



Certificate of Conformity

VWR TISSUE GRINDERS, PLAIN PLUNGER, POTTER-ELVEHJEM TYPE

VWR International certifies that the products comply in every respect with the relevant particular specifications and/or drawings and/or relevant standards and regulations in force, mentioned in this document.

VWR International certifie que les produits sont conformes à tous égards aux spécifications particulières appropriées et / ou dessins, normes et règlements en vigueur, mentionnées dans le présent document

VWR International bestätigt, dass das Produkte allen zutreffenden Spezifikations- und/oder Zeichnungs- und/oder Standardanforderungen sowie den im Dokument aufgeführten gültigen Vorschriften entspricht.

Cat. No.	Capacity (ml)	Total Length(mm)	Pk
432-0206	2	230	1
432-0207	5	235	1
432-0208	10	270	1
432-0209	15	270	1
432-0210	30	270	1
432-0211	50	270	1

Material: PFTE

3M® DYNEON® TF1620 – Compliance Status

FDA – Directive FDA 21 CFR 177.1550

USP Class VI – USP 27, NF 22, 2004 for Plastics at 70°C.

RoHS 2 – Directive 2011/65/EU & Dangerous Substances – 76/769/EC

EU REACH Regulations – SVHC Current Candidate List

Bisphenol A – Absent

Latex – Absent

ADCF – Animal Derived Components Free (BSE/TSE). GMO Free

Endotoxins – Free

POP, PFOS, PFOA, APFO, DEHIP, VOC Free

ODC – Class 1

Dodd-Frank Act – Gold, Coltan, Cassiterite, Wolframite and their derivatives originating from the Democratic Republic of Congo - Absent

Certificate of Conformity

Other characteristics :

- Tissue grinders are used for controlled reduction of particle size and homogenisation of a variety of substances.
- Grinding chamber clearance : 0.15 - 0.25 mm
- Diameter of stainless steel shaft is 6.5 mm.

Disclaimer: VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.

Date of issue: **November 2017**