



Certificate of Conformance and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PSME0001
Lot Number:	1090390-0-0-1
Manufacturing Date:	09/09/2024
Isopropyl Alcohol:	70 ± 3% by volume with the remainder DI H2O.
Expiration Date:	09/2026

Sterility

Gamma Processing Run ID:	3624173
Irradiation Dose Specified:	25.0 kGy to 50.0 kGy
Irradiation Dose Delivered:	29.7 kGy to 41.5 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

QC Supervisor
coc@contecinc.com

Friday, September 20, 2024

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Duplex
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 8621B



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

09/16/2024 14:22:22 GMT
Page 1 of 1

Customer Name: Contec- PO Box 530
P.O.# 144111

Processing Facility: Charlotte Gamma

Work Order # 3624173
Sales Order # 3639302

25-50 kGy

CNT6, Gamma Treatment

Irradiation Date/Time: 09/14/2024 02:36:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	72	CA	PSPS0093	Clean Room Product	1090398-0-0-1	144111
102.000	60	CA	PSME0001	Clean Room Product	1090389-0-0-1	144111
103.000	60	CA	PSME0001	Clean Room Product	1090390-0-0-1	144111
104.000	54	CA	PSPS0044	Clean Room Product	1085645-0-0-1	144111
	246	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	29.7 KGY	Pass	DHOWELL	09/14/2024 10:30:23 GMT
	Reason Code Test					DEREK HOWELL	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	41.5 KGY	Pass	DHOWELL	09/14/2024 10:31:15 GMT
	Reason Code Test					DEREK HOWELL	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG
Reason: Work Order Completions

Date: 09/16/2024 14:20:08 GMT