

Test Report

– Eppendorf PCR clean –

epT.I.P.S.[®] Reloads 2 - 200 µL



10x 96 Eppendorf Tips
in Trays / stacked
PCR clean

LOT K200361Q
1148
2026-11-28

Order no.: 0030 073.819
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 022491733
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: Info@eppendorf.com

• Human DNA

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• PCR inhibition

Method: real-time PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 02.12.21

Testing of the above described lot showed conformity within the limits of detection.

07.12.21

Date of release

Epp 2112-9908

Registration number



Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1