



CERTIFICATE OF MANUFACTURE/ANALYSIS

Product Size:	<u>5 Gallons (18.9 liters)</u>	Product Name:	<u>ProKlenz TWO</u>
Manufacture Date:	<u>September 16, 2019</u>	Product ID:	<u>1422-05WR</u>
Exp. Date:	<u>September 1, 2022</u>	Lot No.:	<u>302603</u>

I. The product listed above is made per the current issue of the following documentation:

Raw Material Quality Control Specifications
Finished Product Master Formula
Manufacturing Instructions
Packaging Specifications
Finished Product Quality Control Specifications

Copies of this documentation are on file at STERIS Corporation, St. Louis Operations and will be provided as required to the applicable regulatory agencies.

II. The product listed above has been analyzed. The release specifications and results are shown below.

<u>Test Name</u>	<u>Method No.</u>	<u>Results</u>	<u>Specification</u>
Color	110	Light Straw	Colorless to Light Straw
Clarity	107	Clear	Clear to Slightly Hazy
pH (1% w/w soln)	106	2.5	2.1 – 2.7
Acidity (as Citric Acid)	5-S	38.1%	36.5 – 39.5%
Sodium Xylene Sulfonate Assay	LC25	0.41%	0.37 – 0.43%
Specific Gravity (@ 25°C)	114	1.181	1.176 – 1.186
Iron	122-S	< 1 ppm	≤ 20 ppm
Mercury	256	< 2 ppb	≤ 8 ppb

Based on these results the product is released for shipment to all STERIS Corporation distribution points.

Signed: *Egison*

Date: September 24, 2019

Quality Representative